

# R740C & R750C Capacitance Transducer Pin-outs

J3	Signal
1	Relay Contact A, N.O.
2	Pressure Output
3	Reserved
4	Reserved
5	Power Return
6	-15 VDC
7	+15V
8	Relay Contact A, N.C.
9	Relay Contact B, N.C.
10	Trip Point A, Setting
11	Trip Point B, Setting
12	Pressure Return
13	Relay contact B, N.O.
14	Relay Contact B, Common
15	Relay contact A, Common